

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1771

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2022 02:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
2:00 196.1 F 3.7  
4:00 215.2 F 8.6  
6:00 224.3 F 8.3  
8:00 234.0 F 9.5  
10:00 238.2 F 9.2  
12:00 245.1 F 13.3  
14:00 244.1 F 12.2  
16:00 257.3 F 19.1  
18:00 268.8 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.8  
2:00 271.5 F 28.1  
3:00 272.6 F 29.0  
4:00 272.0 F 28.4  
  
Min 271.4 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 238.2 F 0.0  
10:00 226.3 F 0.0  
15:00 224.1 F 0.0  
20:00 222.7 F 0.0  
25:00 226.2 F 0.0  
30:00 225.8 F 0.0  
  
Drying Complete

FILLING: 2:08  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:25  
CYCLE COMPLETE

03/30/2022 02:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_