
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002031
8:42 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:44
STERILIZING 3:49
136.5 C 224kPa 3:49
136.4 C 225kPa 4:19
136.4 C 225kPa 4:49
136.4 C 225kPa 5:19
136.3 C 224kPa 5:49
136.4 C 225kPa 6:19
136.3 C 224kPa 6:49
136.3 C 224kPa 7:19
136.3 C 224kPa 7:19
VENTING 7:19
AIR DRYING 8:10
CYCLE COMPLETE 36:16

Digital Signature #
BC3BBA496479A356
