
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002540
10:46 26/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:56
136.5 C 224kPa 3:56
136.3 C 224kPa 4:26
136.4 C 224kPa 4:56
136.1 C 224kPa 5:26
136.3 C 224kPa 5:56
136.4 C 226kPa 6:26
136.3 C 222kPa 6:56
136.4 C 226kPa 7:26
136.4 C 226kPa 7:26
VENTING 7:26
AIR DRYING 8:15
CYCLE COMPLETE 38:15

Digital Signature #
DC8F488E55586425
