
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000525
15:16 28/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:19
STERILIZING 4:32
136.3 C 231kPa 4:32
136.4 C 233kPa 5:02
136.4 C 232kPa 5:32
136.3 C 232kPa 6:02
136.3 C 233kPa 6:32
136.4 C 233kPa 7:02
136.4 C 233kPa 7:32
136.1 C 230kPa 8:02
136.1 C 230kPa 8:02
VENTING 8:02
AIR DRYING 8:57
CYCLE COMPLETE 22:59

Digital Signature #
32C33A93C0981C8B
