
Midmark M11 - v1.0.0
Total Cycles: 653

Name: MIDMARK
Sterilizer ID: V2178198
02/08/2021 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.8 F 0.0
2:00 85.4 F 0.0
4:00 91.3 F 0.1
6:00 127.4 F 0.2
8:00 178.5 F 0.4
10:00 219.4 F 4.3
12:00 232.2 F 7.9
14:00 241.1 F 11.9
16:00 232.7 F 7.2
18:00 245.1 F 12.7
20:00 240.2 F 10.3
22:00 236.4 F 8.5
24:00 251.9 F 15.9
26:00 264.6 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 27.8
1:00 272.6 F 28.5
2:00 271.9 F 28.2
3:00 271.9 F 27.9
4:00 272.7 F 28.6

Min 271.5 F 27.7
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 233.2 F 0.0
10:00 223.1 F 0.0
15:00 226.7 F 0.0
20:00 198.2 F 0.0
25:00 193.8 F 0.0
Dry Cycle Aborted

FILLING: 0:49
HEATING: 27:41
STERILIZING: 4:00
VENTING: 2:55
DRYING: 27:02
TOTAL CYCLE: 01:02:28

02/08/2021 12:35 PM
ACCEPTED: X
REJECTED:
Operator: U01