

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1280

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2021 08:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.8 F 0.0  
2:00 122.1 F 0.0  
4:00 134.0 F 0.1  
6:00 161.0 F 1.8  
8:00 185.3 F 5.1  
10:00 205.4 F 9.0  
12:00 219.0 F 12.8  
14:00 224.7 F 6.3  
16:00 235.3 F 10.6  
18:00 236.7 F 8.1  
20:00 244.1 F 12.6  
22:00 241.2 F 10.8  
24:00 253.1 F 16.7  
26:00 263.5 F 22.8  
28:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.5 F 28.9  
2:00 271.6 F 28.2  
3:00 271.8 F 28.6  
4:00 272.4 F 28.8

Min 271.3 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.6 F 0.0  
5:00 233.1 F 0.0  
10:00 202.9 F 0.0  
15:00 204.6 F 0.0  
20:00 205.3 F 0.0  
25:00 209.2 F 0.0  
30:00 208.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 28:20  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:06:38  
CYCLE COMPLETE

10/20/2021 09:38 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01