
Midmark M11 - v1.0.0
Total Cycles: 4044

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2023 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 178.0 F 3.2
4:00 218.5 F 7.5
6:00 215.3 F 12.5
8:00 223.4 F 7.0
10:00 235.1 F 12.1
12:00 237.1 F 9.1
14:00 239.7 F 9.9
16:00 247.5 F 13.6
18:00 259.1 F 20.0
20:00 270.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.9 F 28.4
2:00 272.3 F 28.8
3:00 271.7 F 28.8
4:00 271.7 F 28.4

Min 271.1 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 201.5 F 0.0
10:00 200.4 F 0.0
15:00 198.5 F 0.0
20:00 194.4 F 0.0
25:00 190.5 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

11/20/2023 05:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____