
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000161
12:43 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:45
PRESSURIZING 0:54
STERILIZING 2:04
136.0 C 224kPa 2:04
136.5 C 224kPa 2:34
136.3 C 225kPa 3:04
136.4 C 224kPa 3:34
136.1 C 223kPa 4:04
136.4 C 225kPa 4:34
136.4 C 224kPa 5:04
136.3 C 224kPa 5:34
136.3 C 224kPa 5:34
VENTING 5:34
AIR DRYING 6:29
CYCLE COMPLETE 6:32

Digital Signature #
7B070F56962B20D1
