
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001082
7:35 12/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:29
STERILIZING 4:34
136.3 C 222kPa 4:34
136.2 C 223kPa 5:04
136.3 C 222kPa 5:34
136.4 C 221kPa 6:04
136.2 C 222kPa 6:34
136.4 C 222kPa 7:04
136.3 C 221kPa 7:34
136.3 C 222kPa 8:04
136.3 C 222kPa 8:04
VENTING 8:04
AIR DRYING 8:56
CYCLE COMPLETE 9:18

Digital Signature #
FE3F79B4679391DA
