

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 004200  
8:51 16/08/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:32  
PRESSURIZING 1:35  
STERILIZING 2:38  
136.4 C 221kPa 2:38  
136.2 C 222kPa 3:08  
136.3 C 224kPa 3:38  
136.4 C 223kPa 4:08  
136.4 C 224kPa 4:38  
136.3 C 222kPa 5:08  
136.5 C 223kPa 5:38  
136.0 C 220kPa 6:08  
136.0 C 220kPa 6:08  
VENTING 6:09  
AIR DRYING 7:00  
CYCLE COMPLETE 8:51  
-----

Digital Signature #  
A39C236036E1EA40  
-----