
Midmark M11 - v1.0.5
Total Cycles: 799

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 166.9 F 2.4
4:00 178.0 F 6.4
6:00 205.1 F 11.8
8:00 216.6 F 7.1
10:00 231.5 F 12.8
12:00 236.1 F 10.5
14:00 235.9 F 9.1
16:00 250.6 F 16.2
18:00 263.0 F 23.3
20:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 273.0 F 30.0
2:00 272.1 F 29.3
3:00 271.9 F 29.6
4:00 272.7 F 30.0

Min 271.4 F 29.1
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 240.8 F 0.0
10:00 234.9 F 0.0
15:00 206.3 F 0.0
20:00 189.3 F 0.0
25:00 205.5 F 0.0
30:00 188.3 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

02/15/2022 10:07 AM
ACCEPTED: X
REJECTED:
Operator: U02