

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3021

Name: RITTER 1  
Sterilizer ID: V2178198  
03/20/2023 02:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.1 F 0.0  
2:00 178.5 F 3.3  
4:00 189.4 F 7.6  
6:00 209.6 F 12.6  
8:00 223.2 F 6.6  
10:00 236.4 F 11.9  
12:00 236.6 F 9.2  
14:00 238.1 F 9.2  
16:00 249.7 F 14.8  
18:00 262.3 F 22.0  
20:00 272.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.4 F 28.4  
2:00 271.7 F 28.2  
3:00 272.5 F 28.5  
4:00 271.4 F 27.9

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
5:00 213.8 F 0.0  
10:00 230.8 F 0.0  
15:00 229.5 F 0.0  
20:00 227.9 F 0.0  
25:00 227.5 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:12  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

03/20/2023 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_