
Midmark M11 - v1.0.0
Total Cycles: 2939

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.5 F 0.0
2:00 160.3 F 2.6
4:00 185.3 F 6.4
6:00 205.2 F 10.6
8:00 220.9 F 4.0
10:00 228.0 F 8.5
12:00 239.9 F 12.4
14:00 238.8 F 10.6
16:00 236.6 F 8.4
18:00 250.2 F 15.3
20:00 261.9 F 22.1
22:00 272.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.3 F 29.2
4:00 272.4 F 28.9

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 231.3 F 0.0
10:00 230.8 F 0.0
15:00 230.2 F 0.0
20:00 229.5 F 0.0
25:00 231.4 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

02/23/2023 02:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____