
Midmark M11 - v1.0.0
Total Cycles: 3836

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2023 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 179.7 F 2.9
4:00 190.8 F 6.9
6:00 217.0 F 11.8
8:00 227.9 F 6.8
10:00 239.6 F 12.0
12:00 236.8 F 9.1
14:00 239.7 F 10.1
16:00 247.2 F 13.5
18:00 258.7 F 19.8
20:00 269.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.5
2:00 272.7 F 28.8
3:00 272.1 F 28.8
4:00 272.0 F 28.4

Min 271.5 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 226.8 F 0.0
10:00 200.1 F 0.0
15:00 195.2 F 0.0
20:00 192.0 F 0.0
25:00 189.0 F 0.0
30:00 186.5 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

09/28/2023 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____