
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008354
15:32 25/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:38
STERILIZING 3:44
136.4 C 224kPa 3:44
136.4 C 226kPa 4:14
136.3 C 220kPa 4:44
136.4 C 224kPa 5:14
136.4 C 224kPa 5:44
136.4 C 224kPa 6:14
136.4 C 224kPa 6:44
136.4 C 224kPa 7:14
136.4 C 224kPa 7:14
VENTING 7:15
AIR DRYING 8:04
CYCLE COMPLETE 8:31

Digital Signature #
E0CC35B6AFD9D2DB
