
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004497
13:38 05/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:24
STERILIZING 3:22
136.4 C 224kPa 3:22
136.3 C 221kPa 3:52
136.2 C 222kPa 4:22
136.4 C 222kPa 4:52
136.4 C 225kPa 5:22
136.3 C 223kPa 5:52
136.4 C 222kPa 6:22
136.4 C 223kPa 6:52
136.4 C 223kPa 6:52
VENTING 6:52
AIR DRYING 7:47
CYCLE COMPLETE 15:02

Digital Signature #
76B408EF7FB86E40
