
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001620
12:47 11/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:28
CYCLE ABORTED 6:11
WATER LEVEL LOW
NOT STERILE
END TIME 6:11

Digital Signature #
2DEE5CBBD468FEFD
