
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000673
12:56 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:07
STERILIZING 4:12
136.4 C 232kPa 4:12
136.1 C 230kPa 4:42
136.4 C 233kPa 5:12
136.2 C 231kPa 5:42
136.4 C 232kPa 6:12
136.2 C 231kPa 6:42
136.4 C 233kPa 7:12
136.5 C 233kPa 7:42
136.5 C 233kPa 7:42
VENTING 7:42
AIR DRYING 8:38
CYCLE COMPLETE 8:39

Digital Signature #
9534AB99E42A23CB
