

-----  
Midmark M11 - v1.0.0  
Total Cycles: 907

Name: RITTER 1  
Sterilizer ID: V2178198  
06/16/2021 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.9 F 0.0  
2:00 167.2 F 1.6  
4:00 174.3 F 4.4  
6:00 202.2 F 8.0  
8:00 211.4 F 12.4  
10:00 223.4 F 7.2  
12:00 233.2 F 12.1  
14:00 234.4 F 9.0  
16:00 242.0 F 10.6  
18:00 244.6 F 12.9  
20:00 256.1 F 18.8  
22:00 266.3 F 25.1  
24:00 272.2 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.8 F 29.4  
2:00 272.0 F 29.4  
3:00 272.2 F 29.0  
4:00 272.8 F 29.5

Min 271.6 F 28.9  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.5 F 0.0  
5:00 249.6 F 0.0  
10:00 228.6 F 0.0  
15:00 230.9 F 0.0  
20:00 225.5 F 0.0  
25:00 224.5 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:03:13  
CYCLE COMPLETE

06/16/2021 03:16 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01