
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000783
16:28 09/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:18
PRESSURIZING 7:14
STERILIZING 8:54
136.0 C 220kPa 8:54
136.2 C 221kPa 9:24
136.0 C 218kPa 9:54
136.3 C 222kPa 10:24
136.2 C 220kPa 10:54
136.4 C 220kPa 11:24
136.0 C 220kPa 11:54
136.0 C 219kPa 12:24
136.0 C 219kPa 12:24
VENTING 12:24
AIR DRYING 13:35
CYCLE COMPLETE 13:40

Digital Signature #
2F1DA840EEDB1DC9
