
Midmark M11 - v1.0.0
Total Cycles: 2168

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 178.3 F 3.8
4:00 194.4 F 8.2
6:00 231.6 F 10.7
8:00 231.0 F 9.1
10:00 235.8 F 9.5
12:00 241.7 F 11.8
14:00 239.0 F 9.9
16:00 252.1 F 16.6
18:00 263.7 F 23.3
20:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.8 F 29.3
2:00 271.8 F 28.7
3:00 271.4 F 28.7
4:00 272.6 F 29.3

Min 271.0 F 28.4
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 238.5 F 0.0
10:00 230.0 F 0.0
15:00 230.6 F 0.0
20:00 230.3 F 0.0
25:00 229.3 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

07/25/2022 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____