
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001082
11:36 18/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:23
STERILIZING 4:37
136.0 C 220kPa 4:37
136.3 C 224kPa 5:07
136.2 C 223kPa 5:37
136.3 C 224kPa 6:07
136.3 C 224kPa 6:37
136.2 C 223kPa 7:07
136.5 C 225kPa 7:37
136.4 C 225kPa 8:07
136.4 C 225kPa 8:07
VENTING 8:08
AIR DRYING 9:00
CYCLE COMPLETE 24:00

Digital Signature #
CC812C2ED7F678E9
