
Midmark M11 - v1.0.0
Total Cycles: 2468

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2022 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 150.9 F 2.2
4:00 170.1 F 5.6
6:00 185.2 F 10.0
8:00 217.1 F 5.3
10:00 225.4 F 7.4
12:00 238.4 F 12.2
14:00 234.7 F 8.9
16:00 240.7 F 10.2
18:00 245.7 F 13.1
20:00 257.9 F 19.7
22:00 268.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.3
2:00 271.9 F 28.6
3:00 271.6 F 28.8
4:00 272.5 F 29.3

Min 271.3 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 231.0 F 0.0
10:00 229.3 F 0.0
15:00 230.2 F 0.0
20:00 225.4 F 0.0
25:00 226.2 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

10/21/2022 10:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____