
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000426
13:11 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:39
136.2 C 223kPa 3:39
136.4 C 233kPa 4:09
136.4 C 233kPa 4:39
136.2 C 233kPa 5:09
136.4 C 233kPa 5:39
136.4 C 233kPa 6:09
136.3 C 232kPa 6:39
136.4 C 233kPa 7:09
136.4 C 233kPa 7:09
VENTING 7:09
AIR DRYING 8:05
CYCLE COMPLETE 8:13

Digital Signature #
6BE0D9F6F9C5455F
