
Midmark M11 - v1.0.0
Total Cycles: 4131

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2023 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.2 F 0.0
2:00 144.9 F 2.5
4:00 162.8 F 5.7
6:00 188.7 F 10.0
8:00 218.9 F 5.0
10:00 228.4 F 8.7
12:00 239.2 F 13.4
14:00 237.5 F 10.1
16:00 235.8 F 8.6
18:00 247.4 F 14.3
20:00 259.5 F 20.8
22:00 270.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.6 F 29.1
2:00 272.3 F 29.5
3:00 271.5 F 29.1
4:00 272.2 F 29.1

Min 270.9 F 28.6
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 225.7 F 0.0
10:00 211.3 F 0.0
15:00 229.7 F 0.0
20:00 220.6 F 0.0
25:00 197.1 F 0.0
30:00 194.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:51
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

12/18/2023 02:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____