

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 008009  
11:59 31/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:29  
STERILIZING 4:33  
136.0 C 220kPa 4:33  
135.8 C 219kPa 5:03  
136.1 C 218kPa 5:33  
136.3 C 222kPa 6:03  
136.4 C 222kPa 6:33  
136.4 C 223kPa 7:03  
136.2 C 220kPa 7:33  
136.4 C 222kPa 8:03  
136.4 C 222kPa 8:03  
VENTING 8:03  
AIR DRYING 8:54  
CYCLE COMPLETE 12:58

-----  
Digital Signature #  
1592EA62C6F2AE8D  
-----