
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000189
14:52 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:47
CYCLE INTERRUPTED
(26)
NOT STERILE
END TIME 5:26

Digital Signature #
4140AEA20C8E94BA
