
Midmark M11 - v1.0.5
Total Cycles: 1658

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 169.3 F 2.4
4:00 168.3 F 5.9
6:00 206.0 F 10.3
8:00 221.9 F 4.9
10:00 227.6 F 10.0
12:00 234.8 F 8.1
14:00 241.0 F 12.5
16:00 239.5 F 10.9
18:00 252.0 F 17.3
20:00 263.5 F 23.9
22:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.5 F 30.0
2:00 271.6 F 29.3
3:00 271.7 F 29.9
4:00 272.1 F 29.8

Min 271.2 F 29.2
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 216.7 F 0.0
10:00 235.4 F 0.0
15:00 231.1 F 0.0
20:00 231.2 F 0.0
25:00 229.2 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

10/26/2022 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____