
Midmark M11 - v1.0.0
Total Cycles: 2791

Name: RITTER 1
Sterilizer ID: V2178198
01/18/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 205.4 F 3.3
4:00 206.7 F 7.4
6:00 214.4 F 13.2
8:00 227.7 F 8.3
10:00 239.5 F 11.6
12:00 240.6 F 11.6
14:00 239.2 F 10.2
16:00 253.1 F 17.1
18:00 265.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.8 F 28.5
3:00 272.2 F 29.2
4:00 272.4 F 29.1

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 239.4 F 0.0
10:00 229.9 F 0.0
15:00 231.5 F 0.0
20:00 222.8 F 0.0
25:00 221.1 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

01/18/2023 11:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____