
Midmark M11 - v1.0.0
Total Cycles: 5286

Name: RITTER 1
Sterilizer ID: V2178198
01/28/2025 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 125.5 F 1.4
4:00 168.0 F 4.1
6:00 168.1 F 7.8
8:00 190.9 F 11.8
10:00 208.2 F 2.6
12:00 218.5 F 6.4
14:00 229.4 F 10.4
16:00 234.4 F 8.3
18:00 240.5 F 11.5
20:00 237.4 F 9.3
22:00 250.0 F 15.4
24:00 260.6 F 21.5
26:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.2
2:00 271.4 F 28.6
3:00 272.3 F 29.3
4:00 271.8 F 28.7

Min 271.3 F 28.6
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 224.1 F 0.0
10:00 232.9 F 0.0
15:00 207.4 F 0.0
20:00 211.2 F 0.0
25:00 206.4 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:49
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:05:28
CYCLE COMPLETE

01/28/2025 02:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____