

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001798  
11:39 19/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 2:18  
STERILIZING 3:20  
136.0 C 220kPa 3:20  
136.2 C 224kPa 3:50  
136.2 C 222kPa 4:20  
136.3 C 225kPa 4:50  
136.1 C 220kPa 5:20  
136.4 C 224kPa 5:50  
136.1 C 223kPa 6:20  
136.3 C 224kPa 6:50  
136.3 C 224kPa 6:50  
VENTING 6:50  
AIR DRYING 7:38  
CYCLE COMPLETE 7:46  
-----

Digital Signature #  
97D23D570980945F  
-----