
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000061
17:14 19/01/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CYCLE ABORTED 4:07
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:07

Digital Signature #
8A680B785BEC7955
