
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000895
17:18 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:51
STERILIZING 5:14
136.4 C 222kPa 5:14
136.4 C 233kPa 5:45
136.4 C 232kPa 6:15
136.3 C 232kPa 6:45
136.4 C 232kPa 7:15
136.4 C 232kPa 7:45
136.4 C 232kPa 8:15
136.4 C 233kPa 8:45
136.4 C 233kPa 8:45
VENTING 8:45
AIR DRYING 9:47
CYCLE COMPLETE 24:47

Digital Signature #
54ACC15CADB6E870
