
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007827
7:34 06/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 4:02
136.4 C 222kPa 4:02
136.1 C 221kPa 4:32
136.4 C 217kPa 5:02
136.1 C 221kPa 5:32
136.4 C 221kPa 6:02
136.3 C 220kPa 6:32
136.4 C 223kPa 7:02
136.3 C 219kPa 7:32
136.3 C 219kPa 7:32
VENTING 7:32
AIR DRYING 8:30
CYCLE COMPLETE 38:30

Digital Signature #
8B0D262019A00506
