

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000264  
15:58 30/09/2020  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 6:05  
STERILIZING 7:15  
136.0 C 229kPa 7:15  
136.4 C 233kPa 7:45  
136.4 C 233kPa 8:15  
136.3 C 233kPa 8:45  
136.4 C 233kPa 9:15  
136.4 C 233kPa 9:45  
136.2 C 233kPa 10:15  
136.4 C 233kPa 10:45  
136.3 C 233kPa 11:15  
136.2 C 231kPa 11:45  
136.3 C 231kPa 12:15  
136.4 C 232kPa 12:45  
136.4 C 233kPa 13:15  
136.3 C 233kPa 13:45  
136.4 C 233kPa 14:15  
136.4 C 233kPa 14:45  
136.3 C 233kPa 15:15  
136.4 C 233kPa 15:45  
136.4 C 231kPa 16:15  
136.4 C 232kPa 16:45  
136.4 C 232kPa 17:15  
136.4 C 232kPa 17:15  
VENTING 17:15  
AIR DRYING 18:23  
CYCLE COMPLETE 32:52  
-----

Digital Signature #  
1067E82F38A85D88  
-----