
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 001019
13:44 25/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:03
PRESSURIZING 1:05
STERILIZING 2:01
136.5 C 233kPa 2:01
136.4 C 233kPa 2:31
136.3 C 230kPa 3:01
136.4 C 230kPa 3:31
136.1 C 230kPa 4:01
136.0 C 230kPa 4:31
136.4 C 233kPa 5:01
136.4 C 233kPa 5:31
136.4 C 233kPa 5:31
VENTING 5:31
AIR DRYING 6:20
CYCLE COMPLETE 6:34

Digital Signature #
FBC472F205D71295
