
Midmark M11 - v1.0.5
Total Cycles: 3936

Name: RITTER2
Sterilizer ID: V2313228
09/12/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 150.8 F 1.6
4:00 158.7 F 4.4
6:00 179.8 F 7.9
8:00 197.8 F 11.5
10:00 215.9 F 4.4
12:00 226.8 F 8.3
14:00 237.1 F 12.5
16:00 235.1 F 9.2
18:00 241.2 F 11.1
20:00 244.9 F 13.0
22:00 255.9 F 18.8
24:00 265.8 F 24.9
26:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.8
2:00 271.6 F 29.3
3:00 273.0 F 29.8
4:00 271.7 F 29.2

Min 271.5 F 29.1
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 226.5 F 0.0
10:00 205.9 F 0.0
15:00 218.4 F 0.0
20:00 174.2 F 0.0
25:00 170.7 F 0.0
30:00 169.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:28
CYCLE COMPLETE

09/12/2024 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____