

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001196  
11:29 28/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:17  
PRESSURIZING 4:18  
STERILIZING 5:25  
136.0 C 220kPa 5:25  
136.4 C 225kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 225kPa 6:55  
136.4 C 223kPa 7:25  
136.3 C 224kPa 7:55  
136.3 C 224kPa 8:25  
136.4 C 225kPa 8:55  
136.4 C 225kPa 8:55  
VENTING 8:55  
AIR DRYING 9:49  
CYCLE COMPLETE 13:55

-----  
Digital Signature #  
EB9165DA431D2480  
-----