
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000118
9:35 11/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:05
PRESSURIZING 4:06
STERILIZING 5:35
136.4 C 220kPa 5:35
136.4 C 224kPa 6:05
136.4 C 223kPa 6:35
136.3 C 223kPa 7:05
136.4 C 221kPa 7:35
136.3 C 222kPa 8:05
136.3 C 222kPa 8:35
136.4 C 221kPa 9:05
136.4 C 221kPa 9:05
VENTING 9:05
AIR DRYING 10:05
CYCLE COMPLETE 10:09

Digital Signature #
297E6EE8E86B933F
