
Midmark M11 - v1.0.0
Total Cycles: 1644

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2022 03:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 184.2 F 3.3
4:00 202.1 F 8.3
6:00 220.2 F 9.7
8:00 230.0 F 8.9
10:00 237.6 F 9.7
12:00 243.6 F 12.5
14:00 242.3 F 11.4
16:00 255.7 F 18.3
18:00 267.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.9
2:00 271.6 F 28.4
3:00 272.7 F 29.0
4:00 271.8 F 28.3

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 231.3 F 0.0
10:00 226.3 F 0.0
15:00 226.1 F 0.0
20:00 231.3 F 0.0
25:00 226.1 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

02/17/2022 07:59 AM
ACCEPTED: X
REJECTED:
Operator: U02