
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001769
13:39 02/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:56
STERILIZING 4:07
136.0 C 221kPa 4:07
136.4 C 222kPa 4:37
136.4 C 226kPa 5:07
136.4 C 225kPa 5:37
136.2 C 224kPa 6:07
136.4 C 224kPa 6:37
136.4 C 225kPa 7:07
136.4 C 225kPa 7:37
136.4 C 225kPa 7:37
VENTING 7:37
AIR DRYING 8:29
CYCLE COMPLETE 8:54

Digital Signature #
9C7988C919F21847
