
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000977
8:50 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:47
136.0 C 224kPa 3:47
136.4 C 233kPa 4:17
136.4 C 233kPa 4:47
136.4 C 232kPa 5:17
136.4 C 232kPa 5:47
136.3 C 231kPa 6:17
136.2 C 230kPa 6:47
136.4 C 233kPa 7:17
136.4 C 233kPa 7:17
VENTING 7:17
AIR DRYING 8:05
CYCLE COMPLETE 10:28

Digital Signature #
0566DD5DA04C43DC
