
Midmark M11 - v1.0.0
Total Cycles: 2336

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 156.4 F 3.1
4:00 169.3 F 6.9
6:00 192.5 F 11.4
8:00 214.0 F 4.7
10:00 225.5 F 9.6
12:00 233.1 F 9.2
14:00 239.6 F 11.5
16:00 237.7 F 9.5
18:00 251.2 F 16.1
20:00 262.8 F 22.8
22:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.1
2:00 271.3 F 28.5
3:00 272.4 F 29.3
4:00 271.7 F 28.6

Min 271.3 F 28.5
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 226.3 F 0.0
10:00 235.7 F 0.0
15:00 231.7 F 0.0
20:00 226.8 F 0.0
25:00 225.5 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

09/19/2022 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____