
Midmark M11 - v1.0.0
Total Cycles: 3881

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2023 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 169.4 F 3.1
4:00 193.9 F 6.9
6:00 209.9 F 12.1
8:00 222.5 F 6.6
10:00 235.3 F 11.5
12:00 234.6 F 8.4
14:00 241.7 F 11.1
16:00 245.5 F 12.9
18:00 257.2 F 19.3
20:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.8
1:00 272.9 F 29.1
2:00 272.8 F 29.4
3:00 271.6 F 28.9
4:00 272.2 F 28.9

Min 271.5 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 191.1 F 0.0
10:00 195.0 F 0.0
15:00 195.2 F 0.0
20:00 194.9 F 0.0
25:00 191.4 F 0.0
30:00 189.8 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 21:53
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

10/10/2023 01:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____