
Midmark M11 - v1.0.5
Total Cycles: 4023

Name: RITTER2
Sterilizer ID: V2313228
10/11/2024 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 137.1 F 1.2
4:00 146.1 F 3.6
6:00 171.7 F 6.7
8:00 190.3 F 10.3
10:00 207.0 F 8.3
12:00 219.1 F 7.4
14:00 228.9 F 11.4
16:00 228.3 F 7.4
18:00 240.3 F 12.7
20:00 239.1 F 10.8
22:00 250.8 F 16.5
24:00 260.9 F 22.3
26:00 270.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.5 F 30.2
2:00 272.2 F 29.8
3:00 271.2 F 29.3
4:00 271.7 F 30.0

Min 271.1 F 29.3
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 210.5 F 0.0
10:00 205.4 F 0.0
15:00 200.7 F 0.0
20:00 201.4 F 0.0
25:00 219.5 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:59
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:43
CYCLE COMPLETE

10/11/2024 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____