
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000946
13:40 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:50
STERILIZING 5:14
136.0 C 227kPa 5:14
136.1 C 230kPa 5:44
136.2 C 230kPa 6:14
136.4 C 233kPa 6:44
136.2 C 231kPa 7:14
136.4 C 232kPa 7:44
136.4 C 232kPa 8:14
136.4 C 232kPa 8:44
136.4 C 232kPa 8:44
VENTING 8:44
AIR DRYING 9:42
CYCLE COMPLETE 24:42

Digital Signature #
AF06DBADE06AE2FD
