
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001433
14:24 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:59
STERILIZING 4:02
136.0 C 220kPa 4:02
136.2 C 224kPa 4:32
136.2 C 224kPa 5:02
136.4 C 226kPa 5:32
136.4 C 226kPa 6:02
136.3 C 224kPa 6:32
136.2 C 224kPa 7:02
136.0 C 223kPa 7:32
136.0 C 223kPa 7:32
VENTING 7:32
AIR DRYING 8:23
CYCLE COMPLETE 13:54

Digital Signature #
2A99EB35E8CB7E48
