
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006675
9:16 01/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:08
136.0 C 220kPa 4:08
135.9 C 220kPa 4:38
136.0 C 219kPa 5:08
135.8 C 218kPa 5:38
136.4 C 219kPa 6:08
136.4 C 224kPa 6:38
136.1 C 218kPa 7:08
136.4 C 223kPa 7:38
136.4 C 223kPa 7:38
VENTING 7:38
AIR DRYING 8:30
CYCLE COMPLETE 8:34

Digital Signature #
951F94AC5C6508F0
