
Midmark M11 - v1.0.5
Total Cycles: 1933

Name: RITTER2
Sterilizer ID: V2313228
01/13/2023 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 77.0 F 0.0
4:00 85.9 F 0.0
6:00 117.3 F 0.0
8:00 151.1 F 0.2
10:00 181.8 F 0.6
12:00 197.6 F 4.1
14:00 214.7 F 7.8
16:00 225.9 F 11.6
18:00 228.1 F 7.0
20:00 239.6 F 11.9
22:00 236.1 F 9.3
24:00 238.8 F 10.0
26:00 246.7 F 13.8
28:00 258.0 F 20.0
30:00 268.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.5
2:00 271.6 F 28.7
3:00 272.2 F 29.6
4:00 272.4 F 29.3

Min 271.2 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 202.2 F 0.0
10:00 248.8 F 0.0
15:00 239.1 F 0.0
20:00 230.0 F 0.0
25:00 229.2 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

01/13/2023 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____