
Midmark M11 - v1.0.0
Total Cycles: 4156

Name: RITTER 1
Sterilizer ID: V2178198
12/27/2023 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 75.4 F 0.3
4:00 85.1 F 1.2
6:00 122.3 F 3.3
8:00 148.6 F 5.7
10:00 173.6 F 9.5
12:00 199.4 F 13.3
14:00 211.8 F 4.0
16:00 221.5 F 7.2
18:00 230.2 F 10.7
20:00 231.5 F 7.6
22:00 239.1 F 10.6
24:00 235.7 F 8.1
26:00 247.2 F 13.6
28:00 258.1 F 19.6
30:00 268.2 F 25.8
32:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.5
2:00 272.7 F 28.9
3:00 271.7 F 28.6
4:00 272.2 F 28.5

Min 271.3 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 253.5 F 0.0
10:00 186.3 F 0.0
15:00 224.1 F 0.0
20:00 217.6 F 0.0
25:00 216.9 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:21
CYCLE COMPLETE

12/27/2023 10:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____